

JUN 26 2009

Supervisor/Operator Performing the Verification Procedure: DataMaster edm S/N 130289 ✓

Name Byron D. Redburn ID# 2087 Date 6-24-09 ✓

A Agency Noome Police Dept Phone # 443-5262

Instrument Location Noome Police Dept ✓

B Alco S/N X301548 ✓ Target Value .079 ✓ High Pressure 700

Alco Test Values .077 ✓ .076 ✓
1st Alco 2nd Alco

Signature [Signature] ✓ 8/1/09

(OVER) [Signature] 7/14/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of Jul, 2009.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

BT9 06/03



Datablaster edn 548 130289

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATABLASTER edn 130289

JUNE 24, 2009

OPERATOR'S NAME:
REDDURN BYRON D
OPERATOR'S NUMBER: 0087
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI:

A
O.L. #: A
DEPT/AGENCY: ONE1
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: X301543

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.91 in
ALCO TARGET .079 06:56
BLANK TEST .000 06:56
INTERNAL STANDARD VERIFIED 06:56
ALCO TV 29.91 in .077 06:57
BLANK TEST .000 06:57
SUBJECT SAMPLE .000 06:58
BLANK TEST .000 06:59
ALCO TV 29.91 in .078 06:59
BLANK TEST .000 07:00

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATABLASTER edn 130289

JUNE 24, 2009
TIME 07:02

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 43c
BREATH TUBE: 42c
BAROMETER: 29.91 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNPOQRSTUVWXYZ[\]^_`abdefghijklmno
pqrstuvwxyz{|}~